

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


MAR 04 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130172 ✓
A	Name	Tim Tuckwood	ID# 3446 Date 2-23-10 ✓
	Agency	Alaska State Troopers	Phone # 895-4868
	Instrument Location	TK AST ✓	
B	Alco S/N	X173017 ✓	Target Value 083 ✓ High Pressure 700 psi ✓
	Alco Test Value Average	083 ✓	084 ✓
		1 st Alco	2 nd Alco
	Signature	Tim Tuckwood ✓	CSB 3/11/10
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of March, 2010.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

